## **Technical Datasheet**

Date: 02.06.22

Page 1 of 1



		EN ISO 13485 certified
Article - No:	04-032-0550	Revision: 4
Remarks	Fitting cap: 04-042-0500 and 04-043-0500	single use only
	Specification	
Product	PCR microcentrifuge tube PP, 0,1ml, without cap, low profile, 8-strips, transparent, np pcr ready max. RCF 20.000g, autocl. up to 121°C	
Colour	transparent	
Nominal Capacity	0,1ml	
Material Characteristics	polypropylene (PP)	
Temperature range	-80°C - 121°C	
Autoclavability	autocl. up to 121°C	
Centrifugation	max. RCF 20.000g	
Product Characteristics	biologically inert	
	physiologically harmless	
	no heavy metals according to EC/94/62	
	Food contact: raw material in compliance with regulation EC 1935/2004	
Purity Grade	free of PCR inhibitors	
	free of human DNA und RNA	
	free of DNase and RNase	
	free of pyrogens	
	free of endotoxins	
Shelf - Life	not limited at controlled storage conditions	
Packaging	pcs per inner cs: 125	
	pcs per carton: 1.250	
	pcs per pallet: 55.000	
Storage	dry condition: 40%-60% rel. humidity	
	bags / cases orig. closed	
	storage temp. +2°C to max. +30°C	
	pal. not stackable	
Approved by nerbe plus	quality department	

nerbe plus GmbH & Co. KG Porschestr. 25 ir D-21423 Winsen/Luhe w Tel: +49 (0)4171 7849 0 H Fax:+49 (0)4171 6 45 45 U

info@nerbe-plus.de www.nerbe-plus.de HRA Lüneburg 203240 US1.-ID: DE 32583 0652 Pers. haftende Gesellschafterin: NERBE R&D GmbH, Winsen/Luhe GF: Philipp J.F. Nerbe HRB Lüneburg 208826 Deutsche Bank AG Kto. 0204 41200 IBAN: DE41 2407 0075 0020 4412 00 SWIFT / BIC DEUTDE2H240 Hamburger Sparkasse Kto. 1387 140831 IBAN: DE56 2005 0550 1387 1408 31 SWIFT / BIC HASPDEHHXXX

